

Certificate of Analysis

Product Information:

Orion Number	940011
Description	94 SERIES ISA 475 ML
Lot Number	LP1
Expiration Date	N/A; Recommendation usage 2 years from date of receipt
Method of Analysis	Potentiometric
Acceptance Range	5.00 M +/- 0.25 M NaNO ₃

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4.90	Moles/Liter Sodium Nitrate

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Nitrate	193	NA

Certificate Date: 11/21/07

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION