

Certificate of Analysis

Product Information:

Orion Number	AC45FZ
Description	4000 NTU FORMAZIN STANDARD, 500ML
Lot Number	LO1
Expiration Date	12/08
Method of Analysis	Turbidimeter
Acceptance Range	4000 +/- 80 NTU Formazin Polymer

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4060	NTU EPA 180

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
NOT APPLICABLE	NA	NA

Certificate Date: 12/21/07

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION