

Certificate of Analysis

Product Information:

Orion Number	011005
Description	COND Std 111.9 mS/cm 5 pk-2oz.
Lot Number	MZ1
Expiration Date	01/2010
Method of Analysis	Conductivity
Acceptance Range	111.9 mS/cm +/- 0.5 mS/cm @ 25C +/- 0.2C

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	112.4	mS/cm at 25 deg. C

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Chloride	2202	NA

Certificate Date: 02/04/08

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION