

Certificate of Analysis

Product Information:

Orion Number	810007
Description	ROSS REF FILL SOLN 3M KCL 5X60ML BOTTLES
Lot Number	MY1
Expiration Date	N/A; Recommendation usage 2 years from date of receipt
Method of Analysis	Potentiometric
Acceptance Range	3.00 M +/- 0.15 M KCL

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	3.04	Moles/Liter Potassium Chloride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 02/21/08

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION