

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	940911
<b>Description</b>	TISAB III 475 ML
<b>Lot Number</b>	MX1
<b>Expiration Date</b>	N/A; Recommendation usage 2 years from date of receipt
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	ppm Fluoride < 0.25 ppm / pH 5.0 +/- 0.2 pH Units

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.13	ppm Fluoride
2	4.8	pH @ 25 degrees C

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA
Dipotassium Hydrogen Phosphate	186-II-g	NA
Potassium Dihydrogen Phosphate	186-I-c	NA
Potassium Hydrogen Phthalate	185H	NA

**Certificate Date:** 03/26/08

**Approved By:**



**Quality Representative**

**Environmental Instruments**  
Water Analysis Instruments

166 Cummings Center

Beverly, MA  
01915

(978) 232-6000  
(978) 232-6015 fax

www.thermo.com/water  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

**Thermo**  
ELECTRON CORPORATION