

Certificate of Analysis

Product Information:

Orion Number	910105
Description	PH 5 BUFFER, 475 ML
Lot Number	NS1
Expiration Date	08/2011
Method of Analysis	Potentiometric
Acceptance Range	5.00 +/- 0.02 pH Units

Actual Analysis:


<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	5.02	pH @ 25 C

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Dihydrogen Phosphate	186-I-g	NA
Disodium Hydrogen Phosphate	186-II-g	NA
Potassium Hydrogen Phthalate	185h	NA

Certificate Date: 08/25/09

Approved By:



Quality Representative

Water Analysis Instruments

North America

166 Cummings Center
Beverly, MA 01915 USA
Toll Free: 1-800-225-1480
Tel: 1-978-232-6000
Dom. Fax: 1-978-232-6015
Int'l Fax: 978-232-6031

Europe

P.O. Box 254, 3860 AG Nijkerk
Wallerstraat 125K, 3862 BN
Nijkerk, Netherlands
Tel: (31) 033-2463887
Fax: (31) 033-2460832

Asia Pacific

Blk 55, Ayer Rajah Crescent
#04-16/24, Singapore 139949
Tel: 65-6778-6876
Fax: 65-6773-0836

www.thermo.com/water

Thermo
SCIENTIFIC