

Certificate of Analysis

Product Information:

Orion Number	CODS01
Description	1000 PPM COD STANDARD
Lot Number	MY1
Expiration Date	02/2010
Method of Analysis	Potentiometric
Acceptance Range	4.167mM +/-0.02mM KHP

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4.170	miliMoles/Liter Potassium Hydrogen Phthalate
2	1002	ppm Chemical Oxygen Demand

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Hydrogen Phthalate	185h	NA

Certificate Date: 02/28/08

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION