

# Certificate of Analysis

## Product Information:

|                           |                          |
|---------------------------|--------------------------|
| <b>Orion Number</b>       | 900022                   |
| <b>Description</b>        | PH CLEANERS B, 4 X 60 ML |
| <b>Lot Number</b>         | MX1                      |
| <b>Expiration Date</b>    | 3/09                     |
| <b>Method of Analysis</b> | Potentiometric           |
| <b>Acceptance Range</b>   | 2.5 % +/- 0.2 Bleach     |

## Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u>      |
|-----------------|----------------|-----------------------------|
| 1               | 2.3            | % Weight by Volume Chlorine |

## Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Potassium Dichromate     | 136e         | NA            |

**Certificate Date:** 04/07/08

**Approved By:**



**Quality Representative**

**Environmental Instruments**  
Water Analysis Instruments

166 Cummings Center

Beverly, MA  
01915

(978) 232-6000  
(978) 232-6015 fax

[www.thermo.com/water](http://www.thermo.com/water)  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

**Thermo**  
ELECTRON CORPORATION