

Certificate of Analysis

Product Information:

Orion Number	977007
Description	STANDARD SOLUTION, 100 PPM CL2, 475ML
Lot Number	MV1
Expiration Date	5/2010
Method of Analysis	Potentiometric
Acceptance Range	0.000475M+/- 0.000003M

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	.000474	Moles/Liter Potassium Iodate

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Dichromate	136e	NA

Certificate Date: 06/03/08

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION