

Certificate of Analysis

Product Information:

Orion Number	841111
Description	NA ISA, 475 ML
Lot Number	MU1
Expiration Date	N/A; Recommendation usage 2 years from date of receipt
Method of Analysis	Potentiometric
Acceptance Range	4.0 +/- 0.1 M Ammonium Chloride - 4.0 +/- 0.1 M Ammonium Hydroxide

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4.06	Moles/Liter Ammonium Chloride
2	4.01	Moles/Liter Ammonium Hydroxide

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Potassium Hydrogen Phthalate	185	h

Certificate Date: 06/13/08

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION