

Certificate of Analysis

Product Information:

Orion Number	040908
Description	FLUORIDE STD, 10 PPM W/ TISAB
Lot Number	LY1
Expiration Date	2/09
Method of Analysis	Potentiometric
Acceptance Range	5.0 ppm +/- 0.10 ppm Fluoride (NaF)

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4.95	ppm Fluoride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA

Certificate Date: 02/20/07

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION