

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FI-01620 Vantaa, Finland

Product Name : Iron

Catalogue no. : 981236

Lot no. : E163

Package : 10 x 20 ml

Storage Conditions : 2...8 °C

Expiry date : 2010-08


### Quality Control

#### Performance test:

<u>Control (lot)</u>	<u>Result (µmol/l)</u>	<u>Target range (µmol/l)</u>
Abtrol (D390)	17.2	14.7-18.7
Nortrol (D389)	53.6	46.3-58.9

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 9/6 2009

  
\_\_\_\_\_  
Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland

## ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy  
Ratastie 2, FI-01620 Vantaa, Finland

Product Name : Iron

Catalogue no. : 981236

Lot no. : E389

Package : 10 x 20 ml

Storage Conditions : 2...8 °C

Expiry date : 2011-01


### Quality Control

#### Performance test:

<u>Control (lot)</u>	<u>Result (µmol/l)</u>	<u>Target range (µmol/l)</u>
Abtrol (D390)	16.7	14.7-18.7
Nortrol (D389)	52.7	46.3-58.9

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 6 / 10 2009

  
Hannele Laine  
Quality Assurance Specialist  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland