



A/S-400

Aluminum with Autoclavable Buckets
8-Place, 50 ml Sealed-
Bucket Assemblies¹

Tube Dimensions:

30 × 115 mm or 17 × 120 mm

Fixed-Angle: 45°

Maximum Speed: 4,380 rpm

Rotor Mass: 2.8 kg

	RCF (g)	R (cm)
Max.	3,009	14.04
Min.	1,209	7.76

For use in T-6000D, T-6000B²,
RT-6000D² and RT-6000B²
centrifuges

Catalog No. 11833	Replacement Parts
Includes: A/S-400 Rotor Sealed-Bucket Assembly, ¹ 50 ml (8 supplied) 11152 O-rings for 50 ml Bucket Caps (1 pkg, 8/pkg) 11827 Vacuum Grease 65937 Instruction Manual 11824	Bucket Cap 11154 Bucket and Cap O-rings (8/pkg) 11827 Sealed-Bucket Assembly ¹ 11152



TUBES, BOTTLES

CAPACITY (ml)		DESCRIPTION	MAX SPEED (rpm)	DIMENSIONS (mm)	TUBE CAT. NO. (Qty/Pkg)	SEALING ASSEMBLY*		ADAPTER REQ'D		NOTES
NOMINAL	FILL					CAT. NO. (Qty/Pkg)	DESCRIPTION	CAT. NO. (Qty/Pkg)	NO. OF PLACES	
50	37	Polypropylene Flanged Tube	6,000	29 × 102	03147 (25)	03268 (25)	Polypropylene, Snap-on	03876 (1)	1	
50	39	Polycarbonate Flanged Tube	6,000	29 × 102	03146 (25)	03268 (25)	Polypropylene, Snap-on	03876 (1)	1	
50	46	Polyallomer Thin-Walled Tube	6,000	28 × 104	03139 (25)	—	—	03876 (1)	1	
50	41	Polypropylene Oak Ridge Bottle	6,000	28 × 103	03530 (25)	03535 (25)	Polypropylene, Sealing	03876 (1)	1	
50	38	Polycarbonate Oak Ridge Bottle	6,000	28 × 103	03480 (25)	03535 (25)	Polypropylene, Sealing	03876 (1)	1	
50	50	Stainless Steel Flanged Tube	6,000	28 × 101	00517 (1)	00518 (1)	Stainless Steel, Sealing	03876 (1)	1	Requires Wrench 01014 for tube sealing.
50	50	Stainless Steel Tube	6,000	28 × 101	00579 (1)	00518 (1)	Stainless Steel, Sealing	03876 (1)	1	Requires Wrench 01014 for tube sealing.
50	—	Tissue Culture Tube, Conical	—	30 × 115	—	—	—	—	—	
35	—	Polyallomer Nalgene® Conical Oak Ridge Tube	—	29 × 105	—	—	—	74234 (6)	—	
35	—	Polycarbonate Nalgene® Conical Oak Ridge Tube	—	29 × 105	—	—	—	74234 (6)	—	
15	12	Corex® Tube	6,000	18 × 102	03286 (25)	—	—	00363 (1) and 03876 (1)	1 1	Requires BOTH adapters.
15	—	Tissue Culture Tube, Conical	—	17 × 120	—	—	—	11148 (1)	1	
12	12	Pyrex® Tube	5,000	18 × 102	03109 (50)	—	—	00363 (1) and 03876 (1)	1 1	Requires BOTH adapters.

A/S-400 rotor also accepts all Tubes, Bottles, and Adapters for SS-34 and SA-600 Rotors, pages B-30 and B-34. Requires Adapter **03876** PLUS any adapter specified in SS-34/SA-600 tube charts.

⁽¹⁾Assembly design tested for microbiological containment (Ref. 1) by PHLS-CAMR, Porton Down. Suitable for use with materials up to ACDP Category 3 (Ref. 2). Seal integrity may be affected by improper use or maintenance.

⁽²⁾Instrument no longer available.

*Sealing assembly parts may be available. See section E for availability.

All trademarks are the property of their respective owners.