

ANALYSIS CERTIFICATE

Manufacturer Name and Address: Thermo Fisher Scientific Oy
Ratastie 2, FIN-01620 Vantaa, Finland

Product Name : MAGNESIUM

Catalogue no. : 981783

Lot no. : D267

Package : 8 x 10 ml

Storage Conditions : 2...25 °C

Expiry date : 2010-05


Quality Control

Performance test:

Control (lot)	Result (mmol/l)	Target range (mmol /l)
Abtrol (B392)	1,23	1,11-1,31
Nortrol (B391)	0,81	0.74-0,86

This product with the mentioned lot number was analysed by our authorised quality control department and confirmed that it meets its specifications.

Vantaa 4 / 6 2008



Heli Kinnunen
QA, Quality Engineer
Quality Regulatory and Compliance
Clinical Diagnostics Finland

Konelab PRIME 60 KUSTI, SN F4620934□
 Konelab User

03/06/2008 09:12


Magnesium 981783, lot D267, exp. 2010-05

Test MG

Accepted - -
 Factor 4.62
 Bias -0.003

Coeff. of det. 0.999948

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Water	-0.003	0.002	0.000	
2	Water	-0.004	-0.002	0.000	
3	sCal	0.219	1.025	1.030	
4	sCal	0.221	1.035	1.030	
5	Nortrol (control)	0.174	0.817	0.800	