

EC DECLARATION OF CONFORMITY

We,

Thermo Fisher Scientific Oy, Clinical Diagnostics Finland

declare that the Chemistry Analyzers with

Trade name:	Aquakem 250	Aquakem 250 Cd
Code:	984200	984201
Type:	974	975

are in conformity with the provisions of the following EC directives, including the latest amendments, and with national legislation implementing these directives:

2006/95/EC:	Low Voltage Directive
2004/108/EC:	EMC Directive

and that the following harmonized standards have been applied:

EN 61010-1
EN 61326-1
EN 61000-6-1
EN 61000-6-3

This Declaration is valid for all instruments which are placed on the market by ourselves on or after May 27th, 2008 and which bear the CE marking.

Vantaa, September 18th, 2008

Thermo Fisher Scientific Oy



Sijja Halme
Quality Manager
Quality, Regulatory and Compliance
Clinical Diagnostics Finland