

## **EC DECLARATION OF CONFORMITY**

We,

**Thermo Fisher Scientific Oy, Clinical Diagnostics Finland**

declare that the Chemistry Analyzers with

Trade name:	<b>Aquakem 200</b>	<b>Aquakem 200 Cd</b>
Code:	<b>984206</b>	<b>984207</b>
Type:	<b>973</b>	<b>981</b>

are in conformity with the provisions of the following EC directives, including the latest amendments, and with national legislation implementing these directives:

2006/95/EC:	Low Voltage Directive
2004/108/EC:	EMC Directive

and that the following harmonized standards have been applied:

EN 61010-1  
EN 61326-1  
EN 61000-6-1  
EN 61000-6-3

This Declaration is valid for all instruments which are placed on the market by ourselves on or after May 27<sup>th</sup>, 2008 and which bear the CE marking.

Vantaa, September 18<sup>th</sup>, 2008

Thermo Fisher Scientific Oy



Silja Halme  
Quality Manager  
Quality, Regulatory and Compliance  
Clinical Diagnostics Finland